

Аналитическая лаборатория ПЭЛ

Meas.No. 347 | Date 11-15-2004 | Time 09:48 | Operator Khotov | ID PEL | Serial No. 0262

ZrO2 LEOTEC
open work

Measuring Range	0.31 [µm] - 300.74 [µm]	Pump	100[rpm]
Resolution	62 Channels (17 mm / 114 mm)	Stirrer	3[rpm]
Absorption	7.00 [%]	Ultrasonic	100
Measurement Duration	3 [Scans]		

Modell Independant

Fraunhofer Calculation selected.

62.47 %	<=	5.00 µm	86.55 %	<=	10.00 µm	97.07 %	<=	15.00 µm
99.80 %	<=	20.00 µm	100.00 %	<=	25.00 µm	100.00 %	<=	30.00 µm
100.00 %	<=	35.00 µm	100.00 %	<=	40.00 µm	100.00 %	<=	45.00 µm
100.00 %	<=	50.00 µm	100.00 %	<=	60.00 µm	100.00 %	<=	70.00 µm
100.00 %	<=	80.00 µm	100.00 %	<=	90.00 µm	100.00 %	<=	100.00 µm

10.00 %	<=	0.60 µm	20.00 %	<=	0.98 µm	30.00 %	<=	1.46 µm
40.00 %	<=	2.15 µm	50.00 %	<=	3.19 µm	60.00 %	<=	4.60 µm
70.00 %	<=	6.30 µm	80.00 %	<=	8.32 µm	90.00 %	<=	11.12 µm
99.99 %	<=	22.14 µm						



